



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Dependable Controls Services, LLC**

**95 Ledge Road, Unit #8  
Seabrook, NH 03874**

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 05 January 2026

Certificate Number: AC-2543



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)**

**Dependable Controls Services, LLC**

95 Ledge Road, Unit #8  
Seabrook, NH 03874  
Dan Snyder (603) 580-5744

**CALIBRATION**

Valid to: **January 5, 2026**

Certificate Number: **AC-2543**

**Chemical Quantities**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters	4 pH 7 pH 10 pH	0.020 pH 0.016 pH 0.039 pH	Comparison to Standard pH Solutions

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouples (Source and Measure) <sup>1</sup>	Type E		Simulation Using or Comparison to Fluke Process Calibrator
	(-200 to 0) °C	0.99 °C	
	(0 to 950) °C	0.82 °C	
	Type J		
	(-200 to 0) °C	1.1 °C	
	(0 to 1 200) °C	0.88 °C	
Type K			
(-200 to 0) °C	1.3 °C		
(0 to 1 370) °C	1.1 °C		

**Chemical Quantities**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouples (Source and Measure) <sup>1</sup>	Type R (-20 to 0) °C	2.8 °C	Simulation Using or Comparison to Fluke Process Calibrator
	(0 to 500) °C	2.2 °C	
	(500 to 1750) °C	1.7 °C	
	Type S (-20 to 0) °C	2.8 °C	
	(0 to 500) °C	2.2 °C	
	(500 to 1 750) °C	1.9 °C	
	Type T (-200 to 0) °C	1.3 °C	
	(0 to 400) °C	0.83 °C	
	Electrical Simulation of RTD's (Source and Measure) <sup>1</sup>	Pt 100 – 385 (-200 to 800) °C	
Pt 100 – 3926 (-200 to 630) °C		0.60 °C	
Pt 100 – 3916 (-200 to 630) °C		0.60 °C	
Pt 200 – 385 (-200 to 250) °C		0.33 °C	
(250 to 630) °C		0.94 °C	
Pt 500 – 385 (-200 to 500) °C		0.49 °C	
(500 to 630) °C		0.61 °C	
Pt 1000 – 385 (-200 to 100) °C		0.29 °C	
(100 to 630) °C		0.54 °C	
Ni 120 (-80 to 260) °C		0.29 °C	

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1</sup>	(0.05 to 1) in	410 μin	Comparison to Gauge Blocks
	(1 to 4) in	630 μin	
	(4 to 6) in	650 μin	
Micrometers <sup>1</sup>	(0.05 to 0.2) in	140 μin	
	(0.2 to 1) in	110 μin	
Indicators <sup>1</sup>	(0.05 to 0.2) in	320 μin	
	(0.2 to 1) in	280 μin	

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(0 to 300) psi	0.14 psi	Comparison to Ashcroft AM2-2 Pressure Transducer
Vacuum <sup>1</sup>	(0 to 28.5) inHg	0.12 inHg	Comparison to Ashcroft AM1 Pressure Transducer
Scales and Balances <sup>1,2</sup>	(1 to 500) mg (0.5 to 10) g (10 to 100) g (100 to 1000) g (1 to 5) kg (5 to 10) kg (10 to 20) kg (20 to 30) kg (30 to 40) kg (40 to 50) kg (50 to 70) kg (70 to 100) kg  50 lb (50 to 100) lb (100 to 200) lb (200 to 300) lb (300 to 400) lb (400 to 500) lb (500 to 600) lb (600 to 700) lb (70 to 800) lb (800 to 900) lb (900 to 1 000) lb	0.40 mg 0.58 mg 2.3 mg 3.4 mg 35 mg 140 mg 210 mg 480 mg 3 100 mg 3 400 mg 6 200 mg 9 200 mg  0.006 5 lb 0.013 lb 0.026 lb 0.039 lb 0.052 lb 0.065 lb 0.078 lb 0.091 lb 0.10 lb 0.12 lb 0.13 lb	Comparison to Class 1, 2, 4, and 6 weights

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Environmental Chambers Humidity	(0 to 54) %RH (55 to 75) %RH (76 to 95) %RH	1.4 %RH 1.6 %RH 2.2 %RH	Comparison to Vaisala HMP75 Thermohygrometer
Temperature	(-10 to 60) °C	0.14 °C	
Temperature Uniformity Survey <sup>1</sup>	(32 to 500) °F (500 to 1 000) °F (1 000 to 1 600) °F (1 600 to 2 100) °F	2 °F 3.1 °F 3.4 °F 3.8 °F	Measurement Using Yokogawa GP20 Temperature Recorder


**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Stopwatches and Timers <sup>1</sup>	5 s to 24 h	0.42 sec	Comparison to Reference Stopwatch
Optical Rotational Speed - Measure	(10 to 100) rpm (100 to 1 000) rpm (1 000 to 10 000) rpm (10 000 to 50 000) rpm	2.1 rpm 2.1 rpm 1.9 rpm 0.016 % reading + 1.1 rpm	Comparison to Non-contact Tachometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The CMC for scales and balances are highly dependent upon the resolution of the unit under test. The uncertainty presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2543.



Jason Stine, Vice President